



LABORATORY REPORT

Example Client, XYZ123
1234 Warde Road
Ann Arbor MI 48108

EXAMPLE, REPORT
WX0000003039 M 12/05/1988 34 Y

Referral Testing

Collected: 04/22/2020 17:22 Received: 04/22/2020 17:22

Table with 6 columns: Test Name, Result, Flag, Ref-Ranges, Units, Site. Contains rows for Hepatitis E Quant by PCR, Source; Hepatitis E Quant by PCR, IU/mL; Hepatitis E Quant by PCR, Log IU/mL; Hepatitis E Quant by PCR, Interp.

INTERPRETIVE INFORMATION: Hepatitis E Virus by Quantitative PCR
The quantitative range of this assay is 3.3- 8.3 log IU/mL (1,800- 180,000,000 IU/mL). One IU/mL of HEV RNA is approximately 2.25 copies/mL.

A negative result (less than 3.3 log IU/mL or less than 1,800 IU/mL) does not rule out the presence of PCR inhibitors in the patient specimen or HEV RNA concentrations below the level of detection of the test. Inhibition may also lead to underestimation of viral quantitation.

Test developed and characteristics determined by ARUP Laboratories. See Compliance Statement B: aruplab.com/CS
Performed By: ARUP Laboratories
500 Chipeta Way
Salt Lake City, UT 84108
Laboratory Director: Julio C. Delgado, MD, MS

Reported Date: 04/22/2020 17:30 HEPCLR
Performing Site:

ARRL: ARUP REFERENCE LAB 500 Chipeta Way Salt Lake City UT 841081221

LAB: L - LOW, H - HIGH, AB - ABNORMAL, C - CRITICAL, . - NOT TESTED