



LABORATORY REPORT

QC ACCOUNT (WARDE)
300 W. TEXTILE
ANN ARBOR MI 48108

EXAMPLE, REPORT W
WX0000003827 M 07/08/1968 56 Y

Molecular

Collected: 03/11/2025 08:26 Received: 03/11/2025 08:26

Table with 6 columns: Test Name, Result, Flag, Ref-Ranges, Units, Site. Row 1: Epstein-Barr DNA PCR, Qualitative, CSF; Row 2: Epstein-Barr Virus DNA, Qualitative, DETECTED, AB, Not detected, WMRL.

This test uses the polymerase chain reaction to amplify regions of the Epstein Barr Virus BLLF1 gene. Real-time detection and quantification are used to determine the viral concentration. The analytical measurement range is 500 to 5 million IU/mL (2.7 to 6.7 log(10) IU/mL). The qualitative limit of detection is 50 IU/mL (1.7 log(10) IU/mL).

Specimens reported as "DETECTED" but <500 IU/mL, contain detectable levels of EB Virus DNA, but the viral load is below the limit of quantification. A "Not Detected" result does not rule out infection.

This test was developed and the performance characteristics determined by Warde Medical Laboratory. It has not been cleared or approved by the FDA. This assay has been validated pursuant to the CLIA regulations and is used for clinical purposes.

Reported Date: 03/11/2025 08:27 EBCQL

Performing Site:

WMRL: WARDE MEDICAL LABORATORY 300 West Textile Road Ann Arbor MI 48108

LAB: L - LOW, H - HIGH, AB - ABNORMAL, C - CRITICAL, . - NOT TESTED

H111000006 Ordered By: KAJAL SITWALA, MD, PHD
WX0000003827 WX000000000002516
Printed D&T: 03/11/25 08:27

Kajal V. Sitwala, MD, PhD - Medical Director
Form: MM RL1
PAGE 1 OF 1